



PMS TIBBİ CİHAZLAR TEKNOLOJİSİ SAN. VE TİC. A.Ş
Karaduvar Mah. Serbest Bölge 11. Cadde
No: 46, Akdeniz, PK: 33020
MERSİN - TURKEY
TEL : (324) 238 70 42 (pbx)
FAX : (324) 238 65 49
E Mail : pms@pmsmedikal.com
www.pmsmedikal.com

EU Directives covered by this Declaration

93/42/EEC MEDICAL DEVICE DIRECTIVE / 2007/47/EC

The Products Covered by this Declaration

*Sterilization Reel
Sterilization Pouch*

EU Representative : Neuster Medizintechnik GmbH Goethestr. 7
40237 Düsseldorf Deutschland Tax ID: 105/5832/2600

The Basis on which Conformity is being Declared

The product identified above complies with the requirements of the Medical Device Directory and its annexes above by meeting the following standard: *EN 868-5, ISO 11140-1 and ISO 11607*

The product is defined as Class 1 product and Declaration of Conformity is issued as per Annex VII of 93/42/EEC MDD.

The technical documentation required to demonstrate that the product meets the requirements of the Medical Device Directive has been compiled by the signatory below and is available for inspection by the relevant enforcement authorities.

The CE mark was first applied in May 2000

The products described above comply with the essential requirements of the directives specified.

Authority: Taner ERSEN

Date : May 2016

PMS PMS TIBBİ CİHAZLAR TEKNOLOJİSİ
SANAYİ VE TİCARET A.Ş.
KARADUVAR MAH. SERBEST BÖLGE 11. CADDE NO:46 33020
ADRES NO:3197829954 MERSİN SERBEST BÖLGE 11
AKDENİZ/MERSİN/TÜRKİYE
T.L.F. 0234 238 70 42 FAX: 0234 238 65 49
Mersis No: 0730038604390024 Ticaret Sicil No: 23486
URUN VERGİ DAİRESİ: 736 030 5015

ATTENTION!

The attention of the specifier, purchaser, installer, or user is drawn to special measures and limitations to use which must be observed when the product is taken into service to maintain compliance with the above directives. Details of these special methods and limitations to use are available on request.

CE DECLARATION OF CONFORMITY